

इंटरनेट

मानक

Disclosure to Promote the Right To Information

Whereas the Parliament of India has set out to provide a practical regime of right to information for citizens to secure access to information under the control of public authorities, in order to promote transparency and accountability in the working of every public authority, and whereas the attached publication of the Bureau of Indian Standards is of particular interest to the public, particularly disadvantaged communities and those engaged in the pursuit of education and knowledge, the attached public safety standard is made available to promote the timely dissemination of this information in an accurate manner to the public.

“जानने का अधिकार, जीने का अधिकार”

Mazdoor Kisan Shakti Sangathan

“The Right to Information, The Right to Live”

“पुराने को छोड़ नये के तरफ”

Jawaharlal Nehru

“Step Out From the Old to the New”

IS 10269 (1982): Haemocytometer Counting Chambers Clinical and Diagnostic Apparatus [MHD 12: Hospital Equipment]



“ज्ञान से एक नये भारत का निर्माण”

Satyanarayan Gangaram Pitroda

“Invent a New India Using Knowledge”



“ज्ञान एक ऐसा खजाना है जो कभी चुराया नहीं जा सकता है”

Bhartrhari—Nitiśatakam

“Knowledge is such a treasure which cannot be stolen”

BLANK PAGE



Indian Standard
SPECIFICATION FOR
HAEMOCYTOMETER COUNTING
CHAMBERS CLINICAL AND
DIAGNOSTIC APPARATUS

UDC 616.155.1—076.3—7



© Copyright 1983

INDIAN STANDARDS INSTITUTION
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 110002

Indian Standard

SPECIFICATION FOR HAEMOCYTOMETER COUNTING CHAMBERS CLINICAL AND DIAGNOSTIC APPARATUS

Medical Glass Instruments and Appliances
Sectional Committee, CPDC 12

Chairman

Representing

MAJ-GEN C. S. V. SUBRAMANYAM Ministry of Defence (DGAFMS), New Delhi

Members

DR K. H. DAVE	Haffkine Bio-Pharmaceutical Corporation Ltd, Bombay
DR A. V. MUTHE (Alternate)	
DR S. G. DEODHARE	Directorate of Medical Education and Research, Bombay
DR I. S. GANDHI	Hindusthan Antibiotics Ltd, Pune
SHRI A. P. JOLLY	All India Instrument Manufacturers & Dealers Association, New Delhi
DR K. G. JOSHI	National Physical Laboratory (CSIR), New Delhi
SHRI P. C. JAIN (Alternate)	
DR K. C. KHERA	Scientific Glass Equipment Manufacturer's Asso- ciation, Bombay
SHRI V. V. KETKAR (Alternate)	
SHRI NARENDRA SINGH	Ministry of Defence (DGI), New Delhi
SHRI J. N. VISHWAKARMA (Alternate)	
SHRI P. R. RAO	Borosil Glass Works Ltd, Bombay
SHRI A. K. BANSAL (Alternate)	
SHRI J. C. SHANDILYA	M/s Top Syringe Mfg Co, Bombay
SHRI J. K. WAD (Alternate)	
DR R. C. SHARMA	Ministry of Health, Directorate General of Health Services, New Delhi
SHRI SARUP SIRCAR	The Scientific Indian Glass Co Ltd, Calcutta
SHRI N. G. SIRCAR (Alternate)	
SHRI R. SOUNDHIRARAJAN	Directorate General of Technical Development, New Delhi
SHRI B. N. DAS (Alternate)	

(Continued on page 2)

© Copyright 1983

INDIAN STANDARDS INSTITUTION

This publication is protected under the *Indian Copyright Act* (XIV of 1957) and reproduction in whole or in part by any means except with written permission of the publisher shall be deemed to be an infringement of copyright under the said Act..

(Continued from page 1 j

Members

DR K. P. SRIVASTAVA

DR A. S. SETHI,
Director (Consr Prod &
Med Instrs)

Representing

The Director, Central Glass & Ceramic Research
Institute, New Delhi
Director General, **ISI** (*Ex-officio Member*)

Secretary

SHRI D. K. AGRAWAL

Deputy Director (Consr Prod & Med Instrs), ISI

Glass Instruments and Appliances Subcommittee, CPDC 12 : 2

Members

COL M. CHAKRABORTY
DR K. H. DAVE

Ministry of Defence (DGAFMS), New Delhi
Haffkine Bio-Pharmaceutical Corporation Ltd,
Bombay

SHRI A. U. MUTHE (*Afternate*)

SHRI HARI DUTTA
LT-COL I. A. KHAN

The **Ganga** Glass Works Pvt Ltd, Balawali
Ministry of Defence (DGI)

SHRI R. P. BHATNAGAR (*Alternate*)

SHRI P. R. RAO

Borosil Glass Works Ltd, Bombay

SHRI **SARUP SIRCAR**

The Scientific Indian Glass Co Ltd, Calcutta

SHRI N. G. SIRCAR (*Alternate*)

Indian Standard
SPECIFICATION FOR
HAEMOCYTOMETER COUNTING
CHAMBERS CLINICAL AND
DIAGNOSTIC APPARATUS

0 . F O R E W O R D

0.1 This Indian Standard was adopted by the Indian Standards Institution on 28 June 1982, after the draft finalized by the Medical Glass Instruments and Appliances Sectional Committee had been approved by the Consumer Products and Medical Instruments Division Council.

0.2 For a long time a need was felt to formulate a national standard for haemocytometer counting chambers. In this standard besides one ruling and two ruling, four ruling chamber has also been included. The four ruling chamber is suitable for *leucocyte* counts, but not for red-cell counts because of drying out of the sample in the fourth and possibly also third ruled area during counting of the first and second areas. The tolerances on dimensions specified in this Indian Standard provide for the highest degree of accuracy which is considered to be attainable in manufacture without unduly increasing the cost of production. Nevertheless, it is important to note that errors as large as ± 7 percent on a blood count could occur if, by chance, all the tolerances permitted by this standard for different parts of the apparatus were cumulative. Therefore, for work where greater accuracy than this is required, it is advisable to determine the accuracy of the apparatus by direct measurement of its dimensions, or other suitable method.

0.3 In the preparation of this standard considerable assistance has been derived from BS 748: 1963 haemocytometer counting chamber and dilution pipettes published by the British Standards Institution.

0.4 For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test, shall be rounded off in accordance with IS : 2-1960*. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

*Rules for rounding off numerical values (revised).

1. SCOPE

1.1 The standard specifies three types of haemocytometer counting chamber with one, two or four rulings and four types of ruling.


2. TYPES OF CHAMBER

2.1 The counting chamber shall consist of a recess in a glass slide bridged by a cover-glass. The base of the recess shall bear either one, two or four rulings of one of the types specified in 8.

2.2 If required by the purchaser the counting chamber may be provided with metalized lines (bright lines), satisfying the requirements specified in 8.4.


3. DESIGN AND DIMENSIONS

3.1 The general design of the glass slides is shown diagrammatically in Fig. 1, 2 and 3.



All dimensions in millimetres.

FIG.1 HAEMOCYTOMETER COUNTING CHAMBER WITH ONE RULING



All dimensions in millimetres.

FIG. 2 HAEMOCYTOMETER COUNTING CHAMBER WITH TWO RULINGS

3.2 The glass slides shall conform to the dimensions given in Table 1. The upper edges of the slides shall be bevelled as shown in Figures 1 to 3. The lower surface of the slide may, if required, have concavity below the ruled area or areas so as to minimize the danger of the lower surface becoming scratched.

TABLE 1 DIMENSIONS OF SLIDES

(Clause 3.2)

TYPE OF CHAMBER	LENGTH mm	BREADTH mm	THICKNESS mm
One ruling	75.0 ± 1.0	25.0 ± 1.0	4.5 to 6.0
Two ruling	75.0 ± 1.0	32.0 ± 1.0	4.5 to 6.0
Four ruling	75.0 ± 1.0	32.0 ± 1.0	4.5 to 6.0

5. LONGITUDINAL GROOVE

5.1 Slides with two or four rulings shall have a longitudinal groove of semi-circular cross-section connecting the centres of the transverse grooves as shown in Fig. 2 and 3.

5.2 The width of this groove measured on the surface of the slide shall be 3.0 ± 0.5 mm.

6. FLATNESS OF UPPER SURFACE OF SLIDE

6.1 When tested conjointly by means of an optical flat, the area of the slide which in use is normally covered by the cover glass shall show not more than three interference fringes in the green light from a mercury vapour lamp.

6.2 The ruled areas between the grooves shall be flat and parallel to the plane of the two outer areas, to within 0.001 mm.

7. DEPTH OF CHAMBER

7.1 For neubauer, improved neubauer and biirker rulings, the mean depth of the areas between the grooves below the plane of the two outer areas shall be 0.100 ± 0.001 mm.


7.2 For modified Fuchs-Rosenthal ruling, the depth shall be 0.200 ± 0.001 mm. except for the four-ruled chamber in which the depth shall be 0.100 ± 0.001 mm.

8. RULINGS

8.1 Position — The ruling shall occupy the position marked 'R' in Figures 1 to 3 and shall be centrally placed between the transverse grooves subject to a tolerance of ± 0.4 mm.


8.2 Type — The ruling shall be one of the following types :

- a) **Neubauer** — This ruling is shown in Fig. 4. Each of the nine large squares into which the ruling is divided has sides 1 mm in length. The central square **ABCD** of Fig. 4, shown on a larger scale in Fig. 5, is subdivided into 16 squares, such as **abcd**, each of which in turn contains 16 squares having sides 0.05 mm in length. Variations are permissible in the ruling of the four corner squares around the central one, but if a ruling of these squares different from that shown in Fig. 4 is required, details shall be specified by the purchaser. Mirror images of the ruling shown in Fig. 4 are permitted.



All dimensions in millimetres.

FIG. 4 NEUBAUER RULING



All dimensions in millimetres.

FIG. 5 CENTRAL PORTION OF NEUBAUER RULING

- b) **Improved Neubauer** — This ruling is shown in Fig. 6. Each of the 9 large squares into which the ruling is divided has sides 1 mm in length. The central square **ABCD** of Fig. 6, shown on a larger scale in Fig. 7, is subdivided into 25 squares such as **abcd**, each of which in turn contains 16 squares having sides 0.05 mm in length.
- c) **Bürker** — This ruling is shown in Fig. 8. Each of the 9 large squares such as **ABCD** into which the ruling is divided has sides 1 mm in length. Each of these large squares contains 16 small squares such as **abcd**, having sides 0.2 mm in length, the subdivision being effected by pairs of lines 0.05 mm apart.
- d) **Modified Fuchs-Rosenthal** — This ruling is shown in Fig. 9. Each of the nine large squares such as **ABCD** into which the ruling is divided has sides 1 mm in length. These squares are indicated by a triple ruling approximately 0.05 mm wide. Each of these large squares is divided into 16 small squares, such as, **abcd**, having sides 0.25 mm in length, the subdivision being effected by single lines.

NOTE — The counting areas used should be areas such as **abcd** for red corpuscles or **ABCD** for white corpuscles, and not the remaining small squares or rectangles. This applies to all rulings except the Modified Fuchs-Rosenthal which is used for counting white corpuscles only.


8.3 Accuracy

- a) In each type of ruling the length of each side, of each square such as **ABCD** shall be 1.000 ± 0.004 mm.
- b) The length of each side of each square such as **abcd** shall be 0.200 ± 0.002 mm in Neubauer, Improved Neubauer and **Bürker** rulings; and shall be 0.250 ± 0.002 mm in Modified **Fuchs-Rosenthal** ruling.
- c) The lines shall be sufficiently fine and uniform in thickness to enable measurements to be taken to an accuracy compatible with the above tolerances.


8.4 Metallized Markings and Films — Any metallized markings or films shall be unaffected by contact with 2 percent acetic acid under normal conditions of use.

9. Normally two cover-glasses are supplied by the manufacturer of the counting chambers. When they are supplied along with the counting chambers, they shall comply with the requirements specified in 9.1 to 9.6.


9.1 The cover-glasses shall be well-annealed and free from any defect which would interfere with clear vision through the glass when viewed through a microscope.



All dimensions in millimetres.
FIG. 6 IMPROVED NEUBAUER RULING



All dimensions in millimetres.
FIG. 7 CENTRAL PORTION OF IMPROVED NEUBAUER RULING



All dimensions in millimetres.

FIG. 8 BÜRKER RULING

9.2 The corners shall be rounded.


9.3 Cover-glasses for slides with one ruling shall be 22 x 16 mm, for those with two rulings 25 x 22 mm and for those with four rulings 35 x 25 mm. The tolerance on each of these dimensions shall be ± 1 mm.

9.4 The mean thickness of a cover-glass for slides with one or two rulings shall be 0.45 ± 0.05 mm, and for slides with four rulings, it shall be 0.55 ± 0.05 mm. The total variation in thickness for any cover-glass shall not exceed 0.025 mm.

9.5 When laid on an optical flat and moved about under gentle pressure for a few seconds the cover glasses shall give not more than 5 interference fringes for the smallest size and 7 fringes for the two larger sizes.

9.6 If dilution pipettes are supplied along with Haemocytometer then pipettes shall conform to IS : 3742-1966*.

*Specification for pipettes, dilution for haemocytometers.



All dimensions in millimetres.

FIG. 9 MODIFIED FUCHS-ROSENTHAL RULING

10. MARKING

10.1 Each slide shall be permanently and legibly marked on the upper surface in accordance with 10.2 to 10.7.

10.2 Either 'Depth 0.1 mm' for slides with Neubauer, Improved Neubauer or **Bürker** rulings or for slides with Modified Fuchs-Rosenthal with four ruled chambers.

Or 'Depth 0.2 mm' for slides with modified Fuchs-Rosenthal rulings with one ruled or two ruled chambers.

10.3 Dimensions of small squares (**abcd**) ruled shall be:

Either '1/400 mm²' for slides with Neubauer or Improved Neubauer rulings.

Or '1/25 mm²' for slides with **Bürker** rulings.

Or '1/16 mm²' for slides with Modified Fuchs-Rosenthal rulings.

10.4 Name of the type of ruling.**10.5** Identification numbers, if required.**10.6** Manufacturer's name or recognized trade-mark.**10.7 ISI Certification Marking** — The counting chambers may also be marked with **ISI** Certification Mark.

NOTE — The use of the **ISI** Certification Mark is governed by 'the provisions of the Indian Standards Institution (Certification Marks) Act and the Rules and Regulations made thereunder. The **ISI** Mark on products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard under a well-defined system of inspection, testing and quality control which is devised and supervised by **ISI** and operated by the producer. **ISI** marked products are also continuously checked by **ISI** for conformity to that standard as a further safeguard. Details of conditions under which a **licence** for the use of the **ISI** Certification Mark may be granted to manufacturers or processors, may be obtained from the Indian Standards Institution.

11. PACKING**11.1** Packing shall be done as agreed to between the manufacturer and the purchaser.

INTERNATIONAL SYSTEM OF UNITS (SI UNITS)

Base **Units**

<i>Quantity</i>	<i>Unit</i>	<i>Symbol</i>
Length	metre	m
Mass	kilogram	kg
Time	second	s
Electric current	ampere	A
Thermodynamic temperature	kelvin	K
Luminous intensity	candela	cd
Amount of substance	mole,	mol

Supplementary Units

<i>Quantity</i>	<i>Unit</i>	<i>Symbol</i>
Plane angle	radian	rad
Solid angle	steradian	sr

Derived **Units**

<i>Quantity</i>	<i>Unit</i>	<i>Symbol</i>	<i>Definition</i>
Force	newton	N	1 N=1kg. m/s ²
Energy	joule	J	1 J=1 N.m
Power	watt	W	1 W=1 J/s
Flux	weber	Wb	1 Wb=1 V.s
Flux density	tesla	T	1 T=1 Wb/m ²
Frequency	hertz	Hz	1 Hz=1 c/s (s ⁻¹)
Electric conductance	sieniens	S	1 S=1 A/V
Electromotive force	volt	V	1 V=1 W/A
Pressure, stress	pascal	Pa	1 Pa=1 N/m ²